

**Office of Chairman Purchase Committee & Dean MGM Medical College Indore**

**Online tender no CPC/EQP-RD/13-14/06**

**SUPPLY OF RADIODIAGNOSIS EQUIPMENTS**

**ONLINE TENDER OPENING (Online & Physical Document submission)**

| <b>Sno</b> | <b>Bidder Name</b>                                   | <b>Status</b>       | <b>Remark</b>  |
|------------|--|---------------------|--|
| 1          | Allengers Medical System, Chandigarh                 | Qualified           | -  |
| 2          | Care stream healthcare Ind. Pvt. Ltd., Mumbai        | Qualified           | -  |
| 3          | Philips Electronics India Ltd, New Delhi             | Qualified           | -  |
| 4          | Siemens Ltd., Mumbai                                 | Qualified           | -  |
| 5          | AGFA Health care Ind. Pvt. Ltd., Thane Mumbai        | Qualified           | -  |
| 6          | Prognosis Medical system Pvt. Ltd., Bangalore        | Qualified           | -  |
| 7          | ADS Diagnostic Ltd.                                  | Qualified           | -  |
| 8          | <b>Fujifilms Ind. Pvt. Ltd., Mumbai</b>              | <b>Disqualified</b> | <b>Online bid not submitted.<br/>Physically document submitted</b> |
| 9          | <b>Cura Healthcare Pvt. Ltd., Tambaram Kadapperi</b> | <b>Disqualified</b> | <b>Online bid not submitted.<br/>Physically document submitted</b> |

**ONLINE TENDER FEES AND EMD (ENVELOPE "A")**

| <b>S.no</b> | <b>Bidder Name</b>                            | <b>Status</b> | <b>Remark</b> |
|-------------|---|---------------|---------------|
| 1           | Allengers Medical System, Chandigarh          | Qualified     | -             |
| 2           | Care stream healthcare Ind. Pvt. Ltd., Mumbai | Qualified     | -             |
| 3           | Philips Electronics India Ltd, New Delhi      | Qualified     | -             |
| 4           | Siemens Ltd., Mumbai                          | Qualified     | -             |
| 5           | AGFA Health care Ind. Pvt. Ltd., Thane Mumbai | Qualified     | -             |
| 6           | Prognosis Medical system Pvt. Ltd., Bangalore | Qualified     | -             |
| 7           | ADS Diagnostic Ltd.                           | Qualified     | -             |

**ONLINE TENDER DOCUMENT SCRUTINY**

| <b>Sno</b> | <b>Bidder Name</b>                            | <b>Status</b> | <b>Remark</b> |
|------------|---|---------------|---------------|
| 1          | Allengers Medical System, Chandigarh          | Qualified     | -             |
| 2          | Care stream healthcare Ind. Pvt. Ltd., Mumbai | Qualified     | -             |
| 3          | Philips Electronics India Ltd, New Delhi      | Qualified     | -             |
| 4          | Siemens Ltd., Mumbai                          | Qualified     | -             |
| 5          | AGFA Health care Ind. Pvt. Ltd., Thane Mumbai | Qualified     | -             |
| 6          | Prognosis Medical system Pvt. Ltd., Bangalore | Qualified     | -             |
| 7          | ADS Diagnostic Ltd.                           | Qualified     | -             |

## TECHNICAL EVALUATION AND DEMONSTRATION ITEMWISE STATUS

Name of Equipment: - Digital Mammography System with Stereotactic Biopsy  
Item Code: - RD 2

### Technical & Demonstration Report

| Sr | Name Of Firm   | Make & Model Quoted  | Qualified/ Disqualified | Remarks  |
|----|----------------|----------------------|-------------------------|--|
| 1  | SIEMENS        | MAMMOMAT INSPIRATION | Qualified               | Demonstrated Equipment meets as per the Tender Specification.  |
| 2  | ALLENERS       | MAM VENUS+           | Disqualified            | <b>All the function quoted was not demonstrated satisfactorily.<br/>No installation of this model exists in the country.<br/>Stereotactic Biopsy not shown.<br/>Undertaking submitted is not acceptable.<br/>Major Deviation</b> |
| 3  | ADS DIAGNOSTIC | LORAD SELENIA        | Qualified               | Demonstrated Equipment meets as per the Tender Specification   |

Name of Equipment: - Digital Radiography System Dual Detector System  
Item Code: - RD 3

### Technical & Demonstration Report

| Sr | Name Of Firm | Make & Model Quoted | Qualified/ Disqualified | Remarks  |
|----|--------------|---------------------|-------------------------|--|
| 1  | ALLENERS     | MARS 80             | Qualifies               | Demonstrated Equipment meets as per the Tender Specification                             |
| 2  | CARESTREAM   | DRX-EVOLUTION       | Qualified               | Demonstrated Equipment meets as per the Tender Specification                             |
| 3  | SIEMENS      | MULTIX FUSION       | Disqualified            | <b>The Auto positioning facility is not available in this model.<br/>Major deviation</b> |
| 4  | AGFA         | DX-D 600            | Qualified               | Demonstrated Equipment meets as per the Tender Specification                             |
| 5  | PROGNOSYS    | PRO RAD 3 NC        | Qualified               | Demonstrated Equipment meets as per the Tender Specification                             |

**Name of Equipment: - Digital Subtraction Angiography Unit (DSA)  
Item Code: - RD 4**

**Technical & Demonstration Report**

| Sr | Name Of Firm | Make & Model Quoted | Qualified/ Disqualified | Remarks  |
|----|--------------|---------------------|-------------------------|--|
| 1  | ALLENGERS    | PRIDE FP            | Qualified               | Demonstrated Equipment meets as per the Tender Specification |
| 2  | PHILIPS      | ALLURA X PER FD 20C | Qualified               | Demonstrated Equipment meets as per the Tender Specification |
| 3  | SIEMENS      | ARTIS ZEE           | Qualified               | Demonstrated Equipment meets as per the Tender Specification |

**Name of Equipment: - Conventional Mammography Machine.  
Item Code: - RD 5**

**Technical & Demonstration Report**

| Sr | Name Of Firm   | Make & Model Quoted | Qualified/ Disqualified | Remarks  |
|----|----------------|---------------------|-------------------------|--|
| 1  | ALLENGERS      | MAM VENUS +         | Qualified               | Demonstrated Equipment meets as per the Tender Specification |
| 2  | ADS DIAGNOSTIC | HOLOGIC LORAD       | Qualified               | Demonstrated Equipment meets as per the Tender Specification |

**Name of Equipment: - Multiloading Computer Radiology System CR System  
Item Code: - RD 6**

**Technical & Demonstration Report**

| Sr | Name Of Firm | Make & Model Quoted | Qualified/ Disqualified | Remarks  |
|----|--------------|---------------------|-------------------------|--|
| 1  | CARESTREAM   | CR MAX              | Qualified               | Demonstrated Equipment meets as per the Tender Specification |
| 2  | AGFA         | DXM-DIGITIZER       | Qualified               | Demonstrated Equipment meets as per the Tender Specification |

**Name of Equipment: - Bed Side Color Doppler  
Item Code: - RD 7**

**Technical & Demonstration Report**

| Sr | Name Of Firm | Make & Model Quoted | Qualified/ Disqualified | Remarks  |
|----|--------------|---------------------|-------------------------|--|
| 1  | PHILIPS      | CX-30               | Qualified               | Demonstrated Equipment meets as per the Tender Specification |
| 2  | SIEMENS      | ACUSON X300         | Qualified               | Demonstrated Equipment meets as per the Tender Specification |

**Name of Equipment: - MRI System (1.5 Tesla)  
Item Code: - RD 9**

**Technical & Demonstration Report**

| Sr | Name Of Firm | Make & Model Quoted | Qualified/ Disqualified | Remarks  |
|----|--------------|---------------------|-------------------------|--|
| 1  | PHILIPS      | ACHIEVA 1.5         | Qualified               | Demonstrated Equipment meets as per the Tender Specification |
| 2  | SIEMENS      | MAGNETOM VERIO 1.5  | Qualified               | Demonstrated Equipment meets as per the Tender Specification |

**Name of Equipment: - MRI System (3.0 Tesla)  
Item Code: - RD 10**

**Technical & Demonstration Report**

| Sr | Name Of Firm | Make & Model Quoted | Qualified/ Disqualified | Remarks  |
|----|--------------|---------------------|-------------------------|--|
| 1  | PHILIPS      | ACHIEVA 3.0         | Qualified               | Demonstrated Equipment meets as per the Tender Specification |
| 2  | SIEMENS      | MAGNETOM VERIO 3.0  | Qualified               | Demonstrated Equipment meets as per the Tender Specification |

**Name of Equipment: -Multislice Spiral CT Scanner on Turn Key Basis 64 Slice  
Item Code: - RD 11**

**Technical & Demonstration Report**

| Sr | Name Of Firm   | Make & Model Quoted                        | Qualified/ Disqualified | Remarks  |
|----|----------------|--|-------------------------|--|
| 1  | PHILIPS        | BRILLIANCE<br>INGENUITY 64                 | Qualified               | Demonstrated Equipment meets as per the Tender Specification   |
| 2  | <b>SIEMENS</b> | <b>SOMATOM<br/>DEFINATION<br/>AS EXCEL</b> | <b>Disqualified</b>     | <b>It has only 32 rows of detectors. Hence not as per tender requirement of 64 rows. Deviation. Company did not quote model with 64 rows of detector which is available with them. Major deviation</b> |

**Name of Equipment: - Multislice Spiral CT Scanner on Turn Key Basis 128 Slice  
Item Code: - RD 12**

**Technical & Demonstration Report**

| Sr | Name Of Firm | Make & Model Quoted            | Qualified/ Disqualified | Remarks  |
|----|--------------|--------------------------------|-------------------------|--|
| 1  | PHILIPS      | BRILLIANCE<br>INGENUITY<br>128 | Qualified               | Demonstrated Equipment meets as per the Tender Specification |
| 2  | SIEMENS      | SOMATOM<br>DEFINATION<br>AS+   | Qualified               | Demonstrated Equipment meets as per the Tender Specification |