

महात्मा गाँधी स्मृति चिकित्सा महाविद्यालय, इन्दौर

पत्र क्रमांक 18560/क्रय/2022

इन्दौर दिनांक 23/11/22

संशोधन - पत्र

(Corrigendum)

चिकित्सा महाविद्यालय की ई - निविदा सूचना क्रमांक 17335-37/पर्चेस/2022 दिनांक 05.11.2022 ई-टेन्डर मध्यप्रदेश ई प्रोक्यूमेंट सिस्टम पोर्टल mptenders पोर्टल पर आमंत्रित की गयी थी। जिसकी टेन्डर आई.डी. 2022_DME_231054_1 है। आमंत्रित ई-टेन्डर की प्री-बिड मीटिंग समय सारणी अनुसार दिनांक 21.11.2022 समय दोपहर 12:30 बजे आयोजित की गयी थी। जिसमें उपस्थित निविदाकारों द्वारा टेन्डर डॉक्यूमेंट सामग्री की संख्या (BOQ) एवं नियम शर्तों पर सुझाव एवं आपत्तियों प्रस्तुत की गयी थी। समिती सदस्यों द्वारा प्राप्त सुझाव एवं आपत्तियों का अवलोकन कर ई-टेन्डर डॉक्यूमेंट एवं सामग्री सूची के स्पेशीफिकेशन (BOQ) में निम्नलिखित संशोधन किया जाता है।

S.No	Tender BOQ (Excel File)	Prescribed as BOQ (Excel File) & Tender Document	Corrigendum संशोधन - पत्र
01	BOQ Sr. NO.670	Aortic bovine pericardial bioprosthetic surgical with a three-stage integrity preservation treatment that enables aldehyde-free dry storage, an expandable Co-Cr band and fluoroscopically visible size markers. Sizes-19mm,21mm,23mm,25mm,27mm	Aortic bovine pericardial bioprosthetic surgical valve sizes 19mm,21mm,23mm,25,27mm
02	BOQ Sr. NO.675	Mitral bovine pericardial bioprosthetic surgical valve with a three-stage integrity preservation treatment that enables aldehyde-free dry storage, a Nitinol wifeform for greater flexibility and fluoroscopically visible Stent posts.	Mitral bovine pericardial bioprosthetic surgical valve. Size 23mm, 25mm,27mm,29mm, 31mm,33mm
03	BOQ Sr. NO.717	717- Aortic Valve Size 17mm to 27mm rotatable, Pyrolytic carbon disc, 80-90 degree opening angle, USFDA / CE Certified	Aortic Valve Size 16mm to 27mm rotatable, Pyrolytic carbon disc, 75-90 degree opening angle, USFDA / CE Certified.
04	BOQ Sr. NO.718	S, disc and frame made of 100% pyrolytic cardon, openong angle of 85 degrees , USFDA / CE certified	Aortic valve size 16 to 27 mm with opening angle of 75 to 90 degree Titanium housing Coated. 360 degree rotation. USFDA / CE certified
05	BOQ Sr. NO.719	Mitral Valve Size 23mm to 33mm rotatable, Pyrolytic carbon disc, 80-90 degree opening angle, USFDA / CE Cetified	Mitral Valve Size 23mm to 33mm rotatable, Pyrolytic carbon disc, 75-90 degree opening angle, USFDA / CE Certified
06	BOQ Sr. NO.720	Mitral Valve Size 25mm to 33mm disc and frame made of 100% pyrolytic cardon, openong angle of 85 degrees, USFDA / CE certified.	Mitral Valve Size 25mm to 33mm disc and frame made of 100% pyrolytic carbon, opening angle of 75-90 degrees, USFDA / CE certified. 360 degree rotation.
07	BOQ Sr. NO.832	832 - 26MM 1/2 Circle Taper Point ,Double arm	26MM 1/2 Circle / 3/8 Circle Taper Point , Double arm.
08	BOQ Sr. NO.862	862 - 8X30"/75CM DA, Oval Pledget 3x6MM, firm pledget,TAPER POINT,1/2 CIRCLE, 16mm	8X30"/75CM DA, Oval Pledget 3x6MM,firm pledget, TAPER POINT,1/2 CIRCLE,16mm, 17mm,18mm
09	BOQ Sr. NO.863	863- 8X30"/75CM DA, Oval Pledget 4.5X6 MM, FIRM pledget, TAPER POINT,1/2 CIRCLE, 25mm	8X30"/75CM DA, Oval Pledget 4.5X6 MM, FIRM pledget, TAPER POINT,1/2 CIRCLE, 25mm,26mm,27mm.
10	BOQ Sr. NO. 350 to 378	Mechanical Detachable Coil for Item No. 350 to 378 ?	Detachable coils (for intracranial aneurysm treatment) with mechanical Detachment system.
11	BOQ Sr. NO. 350 to 389 to 393	Cobalt Chromium made flow diverting stent in item 389 to 393 ?	Flow Diverting stent made of cobalt chromium, Nitinol etc.
12	BOQ Sr. No. 928, 929,259, 260,188,189, 190,191,192, 193, 194, 195 ,196,197,676, 686,820,821	Should European CE & USFDA Approved Certification be considered for Product No. 928,929,259,260,188,189,190,191,192,193, 194, 195, 196,197,676,686,820,821?	The above mentioned products USFDA certification or European CE certification will only be considered.

Smt. Namrata Dev,
Assistant Hospital Administrator

Dean
M.G.M. Medical College
INDORE

Smt. Divya Sharma
Depty. Director (Finance)

Dr. Muneer Ahmad Khan
Assistant Professor

Dr. Arpan Choudhary
Associate Professor

Dr. Sumit Pratap Singh
Associate Professor

Dr. Salil Sakalle
Purchase Officer

Dr. Sumit Shukla
Suptintendent/ Cordinator

Dr. A.D. Bhatnagar
Nodel officer/ President Committee